

For checking the concentration and measuring diluting ratio of cutting oil and washing solutions. Also, for checking the moisture of hydraulic oil !

Feature

- Calibration with water only
- Extremely water resistant (IP 65)
- Light & Compact, 100g
- Revolutionary new feature E.L.I.
- Automatic Temperature Compensation (ATC)
- Measurement in 3 seconds
- Ergonomically designed for one-handed operation
- Convenient storage case
- Battery power indicator
- Designed with a strap holder option







Specifications

Measurement Range	:	0.0 to 70.0
Resolution	:	0.1
Measurement Accuracy	:	±0.2
Measurement Temperature	:	10 to 75°C
Ambient Temperature	:	10 to 40°C (ATC)
Power Supply	:	AAA alkaline battery×2
Battery Life	:	Apporox. 11,000 measurements
International Protection Class	:	IP65 Water Resistant
Dimensions and Weight	:	5.5×3.1×10.9cm, 100g

Application

- Metalworking fluid (water based solution) (Cutting Oil, Grinding Oil, Quenching Oil, Rolling Solution, Wire Stretching Oil etc.)
- Metalworking fluid
- Lubricant oil
- Moisture of hydraulic oil
- Washing solutions



How to use (in the case of cutting oil)

Measurement of Diluting Ratio before use

Periodic Inspections of Concentration

In order to prevent putrefaction, the cutting oil solution concentration should be kept at an optimum level. Periodic (at least once a week) inspections with PAL-102S is highly recommended.

Appropriate filling up

Due to evaporation and contamination of the cutting oil solution, undiluted cutting oil and water need to be added periodically. You can use PAL-102S when adjusting the concentration as original level.

Tips for Measurement with PAL-102S

The concentrations of the cutting oil solution vary from a place to a place in a circulating system.

The concentration near the injection nozzle is most stable (Please ensure your safety when taking the solution from the injection nozzle.

In addition, you can check the function of filter layer if you measure all solutions in the tank, the injection nozzle, and the septic tank.

*PAL-102S can measure 3 type of cutting oils, emulsion, soluble, and chemical.

- *If measurement values are unstable, there are 2 methods that you can try.
- ① Diluting solutions with water. Then, measuring and multipling a displayed value by your diluted ratio.
- ② Leaving the sample on the prism for a little while and measure.

Measuring Method



①Place the sample onto the prism surface.



② Press the START key.



③The measurement value will be displayed in 3 seconds.

Conversion method (in the case of water cased solutions)

E.g.) For an oil solution, with a concentration of 5% (dilution ratio at 20 times), mix 95ml of water to 5ml of undiluted solution. For this example, assume the displayed value, on the PAL-102S, is 4.0. Next, convert the displayed value using the following relationship to determine the conversion factor; Actual concentration+Displayed value, Ex: (5.0+4.0=1.25). Once the conversion factor in known (1.25 in this example), measuring a sample with an unknown concentration is simple. Taking an unknown sample, assume that the measured value is 3.0. Using the simple formula, displayed value×conversion factor (3.0×1.25) the actual concentration of the sampled measured is 3.75.



http://www.atago.net/ overseas@atago.net

- CATAGO U.S.A., Inc. TEL: 1-425-637-2107 CATAGO INDIA Instruments Pvt. Ltd. CATAGO BRASIL Ltda. CATAGO ITALIA s.r.i. CATAGO CHINA Guangzhou Co., Ltd. TEL : 86-20-38108256 ATAGO NIGERIA Scientific Co., Ltd.
 - TEL: 66-21948727-9 TEL : 55 16 3913-8400 TEL : 39 02 36557267 TEL: 7-812-777-96-96 TEL: 234-707-558-1552

2-6-3 Shiba-koen, Minato-ku, Tokyo 105-0011, Japan TEL: 81-3-3431-1943 FAX: 81-3-3431-1945

customerservice@atago-usa.com TEL: 91-22-28544915, 40713232 customerservice@atago-india.com customerservice@atago-thailand.com customerservice@atago-brasil.com customerservice@atago-italia.com info@atago-china.com info@atago-russia.com atagonigeria@atago.net



All ATAGO refractometers are designed and manufactured in Japan.

HACCPGMP GLP ATAGO products comply with HACCP,GMP, and GLP system standards.